



TERA Radon Program

Radon Monitor RADIM 3A
Technical Specifications



v.1 – 2017

Table of Contents

| | | |
|---|-------------------------------|---|
| 1 | Introduction..... | 2 |
| 2 | Use of the Instrument | 2 |
| 3 | Scope of Delivery | 2 |
| 4 | Technical Specifications..... | 3 |
| 5 | Repairs..... | 3 |
| 6 | Warranty..... | 3 |
| 7 | Revision History | 4 |

COPYRIGHT NOTICE

No part of this document may be reproduced, republished, translated or digitalized in any form or by any means, without prior written permission of TESLA.

Information contained in this manual relates exclusively to the TERA system component specified on the title page. New versions and modifications may be developed without prior notice to current users. TESLA has made every attempt to provide you with complete, error-free and accurate information in this manual. TESLA is not liable for errors or omissions contained in this document, or for any damages however resulted from using or relying on any information contained herein. TESLA's liability for errors shall be strictly limited to correcting such errors and providing advisory services as described below.

Users should be familiar with operation basis of used product. If you experience any problems with your product, please contact us at:

TESLA
Podebradska 56/186
180 66 Prague 9
www.tesla.cz

1 Introduction

This document describes technical specifications of the Radon Monitor RADIM 3A.

Product was developed and manufactured in the Czech Republic. All rights reserved TESLA. Offer or delivery of products or services related to the product does not include transfer of ownership rights.

Before using the product, please read this manual carefully and understand all operating and safety precautions. Compliance with operational and safety precaution can prevent from damage to equipment or injuries to personnel. Operating and safety instructions in the document are marked as follows:

Attention! This formatted text indicates the operating and safety instructions.

The product may only be used in the specified manner and for its intended purpose. The product may be provided to third persons along with this documentation only.

2 Use of the Instrument

The Radim3A instrument is developed for monitoring indoor radon concentration, temperature, humidity and pressure.

3 Scope of Delivery

The Radim3A monitor is supplied as a complete unit consisting of the following components:

- the Radim3A instrument with an instruction manual,
- the program for storing and processing data in PC,
- the cable, the line adapter and the transport box.

4 Technical Specifications

| | |
|-------------------------------|---|
| Measured quantity: | air radon concentration |
| Functioning principle: | the radon diffuses into the chamber, covered by felt. The felt absorbs the air-borne radon decay products. The radon activity is determined by measuring the α -activity of ^{218}Po (RaA), collected by the electric field on the surface of the semiconductor detector. |
| Instrument response: | (0.8 counts/h)/(Bq/m ³) |
| Minimum concentration: | activity determined with a statistical error equal to $\pm 20\%$: 30 Bq/m ³ for 1- hour measurement |
| Maximum concentration: | 150 kBq/m ³ for 1-hour measurement |
| Sampling time: | adjustable from 10 minutes to 24 hours |
| Effect of humidity: | the change in the relative humidity from 50% to 90% decreases the sensitivity for -6.5% . The results are automatically corrected for the effect of humidity |
| Electronics: | low power, protection of data |
| Power: | battery powered- LiOn batteries with capacity of 2.5 Ah |
| Current: | 0.49mA during measurement, 6 μA during stand up |
| Operation time: | 210 days from the battery, from main: limited by the memory capacity |
| Memory: | a maximum of 16096 measurements i.e. results of 5060 hours-210 days of monitoring can be stored in memory when 1- hour sampling time is adjusted. |
| Thermometer: | precision of $\pm 0.5\text{ C}$ |
| Manometer: | range from 750 to 1150 hPa, reading differentiation 0.1hPa, precision $\pm 3\text{hPa}$, temperature coefficient $\leq 0.1\text{hPa}/^\circ\text{C}$ |
| Hygrometer: | range from 10 to 95%, precision $\pm 3\%$, the computer determines the absolute humidity (g/m ³) |
| Display: | LCD, 2x 16-characters, |
| Communication: | RS232 standard port |
| Control: | 3 buttons, remote control by PC |
| Dimensions and Weight: | 230 x 170 x 90 mm, cca 2.5 kg |
| Climatic conditions: | temperature 5 to 40 C°, relative humidity 10-95% |
| Calibration: | carried out by manufacturer and attested by The State Metrological Institute of the Czech Republic |

5 Repairs

Any repairs and non basic maintenance must be performed exclusively by TESLA manufacturer.

TESLA
Podebradska 56/186
180 66 Prague 9
www.tesla.cz

6 Warranty

- This product is covered by warranty of 24 months from purchase date.
- In case of warranty claim, please contact our Service Department.
- Warranty covers any defects in materials or workmanship and excludes any damage resulting from or caused by transport or handling or by any misuse.
- Warranty ceases if product has been used improperly or its seal is broken.
- In case of warranty claim, warranty period is prolonged by number of days product was undergoing warranty repairs.
- After the end of its life, product must be handled as e-waste.

7 Revision History

| Revision | Date | Comments |
|-----------------|-------------|-----------------|
| Rev.1: | 31.1.2017 | Initial release |